

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000848  
12:16 07/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:47  
PRESSURIZING 2:55  
STERILIZING 4:06  
136.0 C 228kPa 4:06  
136.3 C 232kPa 4:36  
136.1 C 229kPa 5:06  
136.0 C 229kPa 5:36  
136.4 C 232kPa 6:06  
136.3 C 232kPa 6:36  
136.4 C 233kPa 7:06  
136.2 C 230kPa 7:36  
136.2 C 230kPa 7:36  
VENTING 7:36  
AIR DRYING 8:26  
CYCLE COMPLETE 23:26  
-----

Digital Signature #  
5491870469AFD9CA  
-----