

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2290

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2022 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
2:00 182.3 F 2.5  
4:00 193.2 F 7.1  
6:00 214.3 F 12.5  
8:00 222.7 F 7.8  
10:00 245.0 F 13.3  
12:00 240.1 F 10.5  
14:00 235.9 F 8.4  
16:00 250.9 F 15.3  
18:00 262.8 F 22.2  
20:00 272.4 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 271.8 F 28.3  
2:00 272.4 F 28.5  
3:00 272.5 F 28.8  
4:00 272.5 F 29.0  
  
Min 271.4 F 28.3  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
5:00 220.7 F 0.0  
10:00 197.0 F 0.0  
15:00 193.2 F 0.0  
20:00 190.6 F 0.0  
25:00 217.8 F 0.0  
30:00 233.6 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:53  
CYCLE COMPLETE

09/06/2022 10:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_