

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001419  
12:50 09/09/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:08  
PRESSURIZING 5:34  
STERILIZING 6:57  
136.0 C 220kPa 6:57  
136.1 C 220kPa 7:27  
136.4 C 224kPa 7:57  
136.4 C 217kPa 8:27  
136.1 C 220kPa 8:57  
136.4 C 222kPa 9:27  
136.3 C 223kPa 9:57  
136.4 C 221kPa 10:27  
VENTING 10:27  
AIR DRYING 11:29  
CYCLE COMPLETE 11:57  
-----

Digital Signature #  
1CCD6719ADAE0FEC  
-----