
Midmark M11 - v1.0.5
Total Cycles: 2002

Name: RITTER2
Sterilizer ID: V2313228
02/02/2023 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 164.8 F 3.3
4:00 181.4 F 7.3
6:00 201.5 F 12.7
8:00 222.0 F 7.5
10:00 235.6 F 12.9
12:00 236.6 F 10.6
14:00 235.5 F 9.3
16:00 250.4 F 16.2
18:00 262.3 F 23.2
20:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.5 F 29.9
2:00 271.5 F 29.4
3:00 272.5 F 30.0
4:00 271.8 F 29.4

Min 271.4 F 29.2
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 218.3 F 0.0
10:00 230.8 F 0.0
15:00 222.1 F 0.0
20:00 221.8 F 0.0
25:00 222.5 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

02/02/2023 04:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____