
Midmark M11 - v1.0.5
Total Cycles: 1539

Name: RITTER2
Sterilizer ID: V2313228
09/27/2022 06:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.7 F 0.0
2:00 79.1 F 0.0
4:00 91.6 F 0.0
6:00 143.6 F 0.2
8:00 156.9 F 0.5
10:00 183.9 F 4.0
12:00 208.6 F 7.4
14:00 225.0 F 11.1
16:00 224.6 F 6.2
18:00 235.9 F 10.5
20:00 233.4 F 8.3
22:00 244.3 F 13.1
24:00 242.1 F 11.8
26:00 254.4 F 18.0
28:00 265.2 F 24.5
30:00 271.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.5
1:00 272.2 F 29.3
2:00 272.6 F 29.5
3:00 272.7 F 29.8
4:00 272.4 F 30.0

Min 271.8 F 29.3
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.1 F 0.0
5:00 221.5 F 0.0
10:00 223.4 F 0.0
15:00 236.6 F 0.0
20:00 234.7 F 0.0
25:00 217.5 F 0.0
30:00 211.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 30:35
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:54
CYCLE COMPLETE

09/27/2022 07:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____