
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000523
14:33 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:27
STERILIZING 4:32
136.3 C 222kPa 4:32
136.4 C 225kPa 5:02
136.3 C 223kPa 5:32
136.4 C 225kPa 6:02
136.4 C 226kPa 6:32
136.4 C 224kPa 7:02
136.2 C 222kPa 7:32
136.4 C 221kPa 8:02
136.4 C 221kPa 8:02
VENTING 8:02
AIR DRYING 8:59
CYCLE COMPLETE 9:00

Digital Signature #
B65FB363F5D22165
