

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4573

Name: RITTER 1  
Sterilizer ID: V2178198  
05/22/2024 01:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.8 F 0.0  
2:00 156.6 F 2.5  
4:00 163.0 F 5.8  
6:00 191.6 F 9.9  
8:00 216.1 F 5.3  
10:00 220.1 F 8.2  
12:00 230.6 F 13.0  
14:00 234.6 F 9.6  
16:00 236.9 F 9.2  
18:00 245.7 F 13.7  
20:00 257.6 F 19.9  
22:00 268.3 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.5 F 29.5  
2:00 272.9 F 29.4  
3:00 271.9 F 29.0  
4:00 271.2 F 28.6

Min 271.1 F 28.6  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
5:00 207.9 F 0.0  
10:00 210.4 F 0.0  
15:00 211.8 F 0.0  
20:00 212.1 F 0.0  
25:00 213.0 F 0.0  
30:00 215.7 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:22  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:01:49  
CYCLE COMPLETE

05/22/2024 02:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_