

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3035

Name: RITTER 1  
Sterilizer ID: V2178198  
03/27/2023 08:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
2:00 192.3 F 3.4  
4:00 191.7 F 8.0  
6:00 210.9 F 13.5  
8:00 226.7 F 8.5  
10:00 239.9 F 10.6  
12:00 240.9 F 11.9  
14:00 239.9 F 10.5  
16:00 254.1 F 17.6  
18:00 266.2 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 29.1  
2:00 271.4 F 28.5  
3:00 272.5 F 29.1  
4:00 271.7 F 28.5  
  
Min 271.4 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 216.5 F 0.0  
10:00 212.9 F 0.0  
15:00 238.0 F 0.0  
20:00 235.9 F 0.0  
25:00 224.1 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 19:33  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:21  
CYCLE COMPLETE

03/27/2023 09:44 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_