
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008488
7:50 02/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:53
STERILIZING 4:07
136.0 C 220kPa 4:07
136.4 C 223kPa 4:37
136.4 C 221kPa 5:07
136.3 C 221kPa 5:37
136.2 C 220kPa 6:07
136.3 C 222kPa 6:37
136.4 C 221kPa 7:07
136.4 C 220kPa 7:37
136.4 C 220kPa 7:37
VENTING 7:37
AIR DRYING 8:28
CYCLE COMPLETE 12:40

Digital Signature #
90AE983E7B1DFF1B
