

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3539

Name: RITTER 1  
Sterilizer ID: V2178198  
07/17/2023 05:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.2 F 0.0  
2:00 170.6 F 2.3  
4:00 200.4 F 6.9  
6:00 219.2 F 13.0  
8:00 231.6 F 8.4  
10:00 238.7 F 11.3  
12:00 243.0 F 12.2  
14:00 241.7 F 11.1  
16:00 255.3 F 18.0  
18:00 266.7 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.3 F 28.6  
2:00 271.9 F 28.9  
3:00 271.7 F 28.4  
4:00 273.0 F 29.2  
  
Min 271.0 F 28.2  
Max 273.1 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.0 F 0.0  
5:00 193.5 F 0.0  
10:00 198.0 F 0.0  
15:00 198.0 F 0.0  
20:00 197.7 F 0.0  
25:00 197.3 F 0.0  
30:00 196.2 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 19:27  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:34  
CYCLE COMPLETE

07/17/2023 06:48 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_