
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000801
12:31 17/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:17
PRESSURIZING 7:22
STERILIZING 9:10
136.0 C 220kPa 9:10
136.3 C 223kPa 9:40
136.1 C 220kPa 10:10
135.9 C 219kPa 10:40
136.4 C 222kPa 11:10
136.1 C 221kPa 11:40
136.4 C 219kPa 12:10
136.1 C 220kPa 12:40
136.1 C 220kPa 12:40
VENTING 12:40
AIR DRYING 13:51
CYCLE COMPLETE 23:19

Digital Signature #
2A0C9337B8844D65
