
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002501
12:27 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:42
STERILIZING 3:54
136.4 C 224kPa 3:54
136.2 C 224kPa 4:24
136.3 C 225kPa 4:54
136.3 C 225kPa 5:24
136.2 C 226kPa 5:54
136.2 C 225kPa 6:24
136.0 C 224kPa 6:54
136.3 C 225kPa 7:24
136.3 C 225kPa 7:24
VENTING 7:24
AIR DRYING 8:19
CYCLE COMPLETE 8:39

Digital Signature #
6DAA017244B8A626
