
Midmark M11 - v1.0.0
Total Cycles: 3711

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2023 12:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.1 F 0.0
2:00 151.3 F 2.5
4:00 167.3 F 5.5
6:00 187.2 F 9.1
8:00 203.9 F 13.5
10:00 223.0 F 6.4
12:00 230.2 F 10.8
14:00 230.9 F 7.1
16:00 242.4 F 12.2
18:00 239.7 F 10.1
20:00 251.6 F 15.9
22:00 262.0 F 21.9
24:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.7 F 28.8
2:00 271.7 F 28.9
3:00 271.5 F 28.8
4:00 271.5 F 28.6

Min 271.4 F 28.6
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 228.0 F 0.0
10:00 212.1 F 0.0
15:00 216.0 F 0.0
20:00 201.5 F 0.0
25:00 201.0 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:04:25
CYCLE COMPLETE

08/31/2023 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____