
Midmark M11 - v1.0.0
Total Cycles: 3402

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2023 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.8 F 0.0
2:00 89.2 F 0.0
4:00 118.8 F 0.0
6:00 138.8 F 0.3
8:00 170.2 F 2.4
10:00 186.7 F 5.5
12:00 202.8 F 8.7
14:00 217.5 F 12.1
16:00 221.8 F 5.7
18:00 232.1 F 9.9
20:00 237.5 F 9.9
22:00 240.8 F 11.3
24:00 237.3 F 9.1
26:00 250.6 F 15.5
28:00 261.6 F 21.7
30:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.6 F 29.1
2:00 271.5 F 28.3
3:00 272.3 F 28.7
4:00 272.0 F 28.8

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 216.3 F 0.0
10:00 219.6 F 0.0
15:00 224.5 F 0.0
20:00 228.6 F 0.0
25:00 220.6 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 30:29
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:17
CYCLE COMPLETE

06/19/2023 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____