

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4440

Name: RITTER2  
Sterilizer ID: V2313228  
03/21/2025 08:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 76.5 F 0.2  
4:00 120.2 F 0.5  
6:00 111.5 F 2.6  
8:00 148.0 F 5.9  
10:00 170.2 F 9.7  
12:00 188.2 F 10.8  
14:00 210.1 F 4.2  
16:00 220.3 F 7.5  
18:00 230.6 F 11.3  
20:00 229.5 F 7.4  
22:00 242.0 F 12.8  
24:00 240.6 F 11.2  
26:00 253.0 F 17.5  
28:00 263.8 F 23.9  
30:00 272.0 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 272.7 F 29.6  
2:00 272.3 F 29.7  
3:00 272.4 F 29.3  
4:00 272.7 F 30.0

Min 271.7 F 29.0  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.1 F 0.0  
5:00 193.6 F 0.0  
10:00 221.4 F 0.0  
15:00 223.2 F 0.0  
20:00 222.3 F 0.0  
25:00 219.9 F 0.0  
30:00 220.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 30:07  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:08:21  
CYCLE COMPLETE

03/21/2025 09:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_