
Midmark M11 - v1.0.0
Total Cycles: 1887

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2022 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.5 F 0.0
2:00 197.9 F 3.6
4:00 201.2 F 8.2
6:00 216.1 F 8.6
8:00 230.3 F 8.9
10:00 240.8 F 10.6
12:00 241.7 F 11.9
14:00 240.3 F 10.4
16:00 253.7 F 17.1
18:00 265.5 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.4 F 28.1
3:00 272.7 F 29.1
4:00 272.0 F 28.4

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 225.8 F 0.0
10:00 228.9 F 0.0
15:00 229.7 F 0.0
20:00 227.4 F 0.0
25:00 225.6 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:42
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

04/28/2022 03:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____