

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3124

Name: RITTER2  
Sterilizer ID: V2313228  
11/14/2023 03:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.9 F 0.0  
2:00 106.8 F 0.4  
4:00 135.2 F 2.4  
6:00 161.9 F 5.9  
8:00 184.7 F 10.3  
10:00 207.8 F 5.1  
12:00 217.3 F 7.8  
14:00 228.4 F 12.5  
16:00 233.5 F 9.1  
18:00 239.4 F 10.3  
20:00 245.5 F 13.4  
22:00 257.5 F 19.9  
24:00 268.4 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.5 F 29.6  
2:00 271.6 F 28.8  
3:00 272.4 F 29.7  
4:00 272.4 F 29.2

Min 271.3 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 221.3 F 0.0  
10:00 199.5 F 0.0  
15:00 193.4 F 0.0  
20:00 191.8 F 0.0  
25:00 190.2 F 0.0  
30:00 189.0 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:14  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:03:25  
CYCLE COMPLETE

11/14/2023 04:35 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_