

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5173

Name: RITTER 1  
Sterilizer ID: V2178198  
12/10/2024 09:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.5 F 0.0  
2:00 69.9 F 0.0  
4:00 79.7 F 0.0  
6:00 111.5 F 0.0  
8:00 142.5 F 0.2  
10:00 172.6 F 0.4  
12:00 196.8 F 0.6  
14:00 211.8 F 2.4  
16:00 222.4 F 5.6  
18:00 230.4 F 8.8  
20:00 239.6 F 12.7  
22:00 236.9 F 9.3  
24:00 240.2 F 10.0  
26:00 246.5 F 13.3  
28:00 244.0 F 12.0  
30:00 256.6 F 18.6  
32:00 267.6 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.0  
1:00 272.2 F 28.2  
2:00 272.7 F 28.6  
3:00 272.5 F 28.7  
4:00 271.7 F 28.0  
  
Min 271.7 F 27.9  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.5 F 0.0  
5:00 199.2 F 0.0  
10:00 191.4 F 0.0  
15:00 192.9 F 0.0  
20:00 226.9 F 0.0  
25:00 235.1 F 0.0  
30:00 228.5 F 0.0  
  
Drying Complete

FILLING: 0:41  
HEATING: 33:19  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:11:24  
CYCLE COMPLETE

12/10/2024 10:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_