

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007039  
13:28 06/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:34  
STERILIZING 3:24  
136.0 C 220kPa 3:24  
136.4 C 222kPa 3:54  
136.3 C 223kPa 4:24  
136.4 C 221kPa 4:54  
136.1 C 217kPa 5:24  
136.3 C 224kPa 5:54  
136.4 C 224kPa 6:24  
136.2 C 222kPa 6:54  
136.2 C 222kPa 6:54  
VENTING 6:54  
AIR DRYING 7:47  
CYCLE COMPLETE 24:33

-----  
Digital Signature #  
672A9AFB5A3CC542  
-----