
Midmark M11 - v1.0.5
Total Cycles: 2163

Name: RITTER2
Sterilizer ID: V2313228
03/20/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.5 F 0.0
2:00 133.5 F 1.8
4:00 153.9 F 4.7
6:00 179.2 F 9.1
8:00 204.2 F 8.8
10:00 223.8 F 8.1
12:00 238.9 F 13.3
14:00 236.3 F 10.4
16:00 234.8 F 8.5
18:00 249.2 F 15.4
20:00 261.1 F 22.3
22:00 271.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.4 F 29.6
2:00 272.6 F 30.0
3:00 271.8 F 29.7
4:00 272.4 F 29.7

Min 271.6 F 29.4
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 230.1 F 0.0
10:00 223.2 F 0.0
15:00 201.3 F 0.0
20:00 219.6 F 0.0
25:00 219.1 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

03/20/2023 02:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____