
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003221
15:23 24/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:04
STERILIZING 4:06
136.4 C 223kPa 4:06
136.4 C 224kPa 4:36
136.4 C 224kPa 5:06
136.3 C 224kPa 5:36
136.3 C 224kPa 6:06
136.4 C 224kPa 6:36
136.4 C 223kPa 7:06
136.1 C 222kPa 7:36
136.1 C 222kPa 7:36
VENTING 7:36
AIR DRYING 8:35
CYCLE COMPLETE 8:38

Digital Signature #
6AC07AF89C264A44
