
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000070
10:26 24/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:54
STERILIZING 4:56
136.0 C 220kPa 4:56
136.3 C 222kPa 5:26
136.4 C 225kPa 5:56
136.3 C 224kPa 6:26
136.0 C 220kPa 6:56
136.4 C 221kPa 7:26
136.4 C 225kPa 7:56
136.0 C 223kPa 8:26
136.0 C 223kPa 8:26
VENTING 8:26
AIR DRYING 9:17
CYCLE COMPLETE 19:19

Digital Signature #
373F4E28727245D4
