
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001444
10:08 09/17/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:10
PRESSURIZING 7:00
STERILIZING 8:44
136.0 C 220kPa 8:44
136.4 C 223kPa 9:14
136.1 C 220kPa 9:44
136.3 C 221kPa 10:14
135.9 C 220kPa 10:44
136.2 C 221kPa 11:14
136.5 C 223kPa 11:44
136.0 C 219kPa 12:14
VENTING 12:14
AIR DRYING 13:20
CYCLE COMPLETE 43:20

Digital Signature #
51B8AA0C73BCC9DB
