
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002640
13:34 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:17
STERILIZING 3:25
136.0 C 220kPa 3:25
136.4 C 226kPa 3:55
136.3 C 224kPa 4:25
136.1 C 222kPa 4:55
136.4 C 225kPa 5:25
136.1 C 223kPa 5:55
136.4 C 222kPa 6:25
136.3 C 225kPa 6:55
136.3 C 225kPa 6:55
VENTING 6:55
AIR DRYING 7:46
CYCLE COMPLETE 14:45

Digital Signature #
BF6DD01643D13F1C
