
Midmark M11 - v1.0.5
Total Cycles: 1834

Name: RITTER2
Sterilizer ID: V2313228
12/13/2022 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 189.3 F 3.6
4:00 196.5 F 9.2
6:00 219.2 F 7.0
8:00 232.2 F 10.8
10:00 233.7 F 8.8
12:00 239.1 F 9.9
14:00 249.0 F 15.1
16:00 261.6 F 22.3
18:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.5 F 29.5
2:00 271.7 F 28.8
3:00 272.6 F 29.7
4:00 272.2 F 29.1

Min 271.6 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 215.9 F 0.0
10:00 224.0 F 0.0
15:00 229.6 F 0.0
20:00 218.3 F 0.0
25:00 216.7 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 2:53
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

12/13/2022 04:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____