

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1788

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2022 05:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
2:00 184.2 F 1.1  
4:00 201.8 F 6.3  
6:00 221.2 F 12.0  
8:00 229.4 F 8.3  
10:00 238.8 F 11.1  
12:00 243.8 F 13.0  
14:00 244.5 F 12.8  
16:00 258.4 F 20.2  
18:00 270.5 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.2  
2:00 271.7 F 29.1  
3:00 272.4 F 29.2  
4:00 271.4 F 28.7  
  
Min 271.4 F 28.6  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 222.9 F 0.0  
10:00 240.7 F 0.0  
15:00 230.4 F 0.0  
20:00 229.5 F 0.0  
25:00 227.2 F 0.0  
30:00 225.3 F 0.0  
  
Drying Complete

FILLING: 1:34  
HEATING: 18:47  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:33  
CYCLE COMPLETE

11/29/2022 06:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_