
Midmark M11 - v1.0.5
Total Cycles: 1804

Name: RITTER2
Sterilizer ID: V2313228
12/05/2022 04:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.4 F 0.0
2:00 166.9 F 3.5
4:00 179.8 F 7.5
6:00 200.6 F 12.7
8:00 218.8 F 7.2
10:00 234.1 F 12.7
12:00 236.0 F 10.3
14:00 236.0 F 8.8
16:00 250.3 F 15.9
18:00 261.9 F 22.9
20:00 273.1 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.4 F 30.0
2:00 272.3 F 29.5
3:00 271.1 F 29.0
4:00 271.8 F 29.9

Min 271.1 F 29.0
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 236.1 F 0.0
10:00 225.8 F 0.0
15:00 221.5 F 0.0
20:00 220.2 F 0.0
25:00 219.9 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

12/05/2022 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____