

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2335

Name: RITTER 1  
Sterilizer ID: V2178198  
09/19/2022 10:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.4 F 0.0  
2:00 176.5 F 3.3  
4:00 182.7 F 7.7  
6:00 211.1 F 12.9  
8:00 223.2 F 7.3  
10:00 237.9 F 12.9  
12:00 238.3 F 10.3  
14:00 237.0 F 8.4  
16:00 250.7 F 15.6  
18:00 262.6 F 22.6  
20:00 272.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.5 F 29.3  
2:00 271.8 F 28.6  
3:00 271.7 F 28.9  
4:00 272.5 F 29.3  
  
Min 271.3 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.6 F 0.0  
5:00 225.0 F 0.0  
10:00 235.4 F 0.0  
15:00 229.2 F 0.0  
20:00 228.5 F 0.0  
25:00 229.2 F 0.0  
30:00 230.2 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:06  
CYCLE COMPLETE

09/19/2022 11:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_