

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3868

Name: RITTER 1  
Sterilizer ID: V2178198  
10/05/2023 07:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.3 F 0.0  
2:00 75.2 F 0.0  
4:00 93.4 F 0.0  
6:00 120.8 F 0.0  
8:00 153.9 F 0.5  
10:00 176.7 F 3.6  
12:00 193.0 F 7.0  
14:00 206.9 F 10.2  
16:00 220.8 F 10.4  
18:00 224.3 F 6.9  
20:00 233.8 F 10.9  
22:00 233.3 F 7.7  
24:00 243.5 F 12.4  
26:00 240.9 F 10.6  
28:00 252.8 F 16.6  
30:00 263.3 F 22.7  
32:00 272.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.1  
1:00 272.6 F 28.9  
2:00 271.4 F 28.2  
3:00 272.3 F 28.6  
4:00 271.9 F 28.6  
  
Min 271.3 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
5:00 211.4 F 0.0  
10:00 205.9 F 0.0  
15:00 219.3 F 0.0  
20:00 209.4 F 0.0  
25:00 210.2 F 0.0  
30:00 208.6 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:09  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:10:24  
CYCLE COMPLETE

10/05/2023 08:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_