
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001703
14:20 07/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:16
STERILIZING 3:19
136.0 C 220kPa 3:19
136.2 C 223kPa 3:49
136.4 C 225kPa 4:19
136.2 C 224kPa 4:49
136.4 C 226kPa 5:19
136.4 C 225kPa 5:49
136.3 C 224kPa 6:19
136.1 C 223kPa 6:49
136.1 C 223kPa 6:49
VENTING 6:49
AIR DRYING 7:42
CYCLE COMPLETE 7:57

Digital Signature #
893957729CF4E851
