
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007486
8:52 04/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:38
STERILIZING 4:46
136.5 C 224kPa 4:46
136.4 C 222kPa 5:17
136.4 C 223kPa 5:47
136.4 C 224kPa 6:17
136.0 C 221kPa 6:47
136.2 C 224kPa 7:17
136.1 C 221kPa 7:47
136.3 C 224kPa 8:17
136.3 C 224kPa 8:17
VENTING 8:17
AIR DRYING 9:16
CYCLE COMPLETE 9:31

Digital Signature #
4E7E5F57C14E5BA1
