
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000148
11:43 14/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 4:30
STERILIZING 5:37
136.0 C 221kPa 5:37
136.4 C 226kPa 6:07
136.1 C 224kPa 6:37
136.4 C 226kPa 7:07
136.4 C 226kPa 7:37
136.4 C 224kPa 8:07
136.2 C 224kPa 8:37
136.4 C 225kPa 9:07
136.4 C 225kPa 9:07
VENTING 9:07
AIR DRYING 10:08
CYCLE COMPLETE 12:04

Digital Signature #
1BDBE9ABCB2703EB
