
Midmark M11 - v1.0.0
Total Cycles: 3564

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2023 05:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.0 F 0.0
2:00 118.8 F 0.0
4:00 133.5 F 0.6
6:00 169.3 F 2.5
8:00 200.4 F 5.2
10:00 202.4 F 8.8
12:00 216.0 F 12.6
14:00 223.6 F 6.6
16:00 234.5 F 10.9
18:00 231.8 F 7.3
20:00 243.9 F 13.1
22:00 242.5 F 11.7
24:00 254.4 F 17.8
26:00 265.0 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.1 F 28.9
2:00 271.6 F 28.9
3:00 272.0 F 28.8
4:00 271.5 F 28.9

Min 271.2 F 28.4
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 208.6 F 0.0
10:00 223.4 F 0.0
15:00 232.6 F 0.0
20:00 233.4 F 0.0
25:00 224.8 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:56
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:06:28
CYCLE COMPLETE

07/21/2023 06:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____