
Midmark M11 - v1.0.5
Total Cycles: 2555

Name: RITTER2
Sterilizer ID: V2313228
06/23/2023 07:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 80.3 F 0.0
4:00 84.0 F 0.0
6:00 112.1 F 0.0
8:00 145.6 F 2.2
10:00 175.4 F 5.3
12:00 197.1 F 8.9
14:00 211.0 F 12.8
16:00 223.1 F 5.7
18:00 230.0 F 9.6
20:00 237.0 F 11.2
22:00 238.5 F 10.5
24:00 236.1 F 8.5
26:00 247.5 F 14.3
28:00 258.3 F 20.3
30:00 268.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.1
1:00 272.8 F 29.8
2:00 271.7 F 28.9
3:00 272.7 F 29.8
4:00 271.6 F 29.0

Min 271.4 F 28.9
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 228.7 F 0.0
10:00 217.5 F 0.0
15:00 222.1 F 0.0
20:00 202.5 F 0.0
25:00 216.4 F 0.0
30:00 212.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:56
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:52
CYCLE COMPLETE

06/23/2023 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____