
Midmark M11 - v1.0.5
Total Cycles: 848

Name: RITTER2
Sterilizer ID: V2313228
03/01/2022 07:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.3 F 0.0
2:00 76.3 F 0.0
4:00 81.3 F 0.0
6:00 108.8 F 0.8
8:00 142.7 F 3.5
10:00 167.4 F 6.3
12:00 188.5 F 9.8
14:00 204.2 F 10.8
16:00 223.7 F 6.8
18:00 235.4 F 11.4
20:00 233.0 F 8.2
22:00 242.4 F 12.0
24:00 244.4 F 12.7
26:00 256.5 F 19.1
28:00 267.2 F 25.8
30:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.7 F 29.5
2:00 271.7 F 29.3
3:00 272.8 F 29.5
4:00 271.7 F 29.2

Min 271.5 F 28.9
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.6 F 0.0
5:00 223.8 F 0.0
10:00 209.5 F 0.0
15:00 178.3 F 0.0
20:00 172.7 F 0.0
25:00 173.2 F 0.0
30:00 156.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 30:13
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:15
CYCLE COMPLETE

03/01/2022 09:08 AM
ACCEPTED: X
REJECTED:
Operator: U02