
Midmark M11 - v1.0.0
Total Cycles: 5107

Name: RITTER 1
Sterilizer ID: V2178198
11/19/2024 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 70.7 F 0.0
4:00 77.4 F 0.0
6:00 105.3 F 0.1
8:00 142.3 F 0.2
10:00 171.5 F 0.6
12:00 196.5 F 1.6
14:00 210.7 F 4.9
16:00 220.0 F 7.9
18:00 229.8 F 11.4
20:00 226.3 F 5.8
22:00 239.0 F 11.0
24:00 233.6 F 7.7
26:00 244.2 F 11.8
28:00 245.4 F 12.6
30:00 257.4 F 18.9
32:00 268.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.7 F 28.6
2:00 271.7 F 27.8
3:00 272.3 F 28.5
4:00 272.5 F 28.4

Min 271.4 F 27.7
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 199.1 F 0.0
10:00 217.8 F 0.0
15:00 225.4 F 0.0
20:00 238.6 F 0.0
25:00 213.7 F 0.0
30:00 210.5 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 33:23
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:11:30
CYCLE COMPLETE

11/19/2024 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____