
Midmark M11 - v1.0.5
Total Cycles: 2096

Name: RITTER2
Sterilizer ID: V2313228
02/27/2023 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.6 F 0.0
2:00 188.5 F 4.3
4:00 203.8 F 9.5
6:00 218.4 F 6.9
8:00 229.8 F 10.9
10:00 233.4 F 8.8
12:00 238.3 F 9.8
14:00 248.9 F 15.4
16:00 261.9 F 22.6
18:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.1
1:00 272.5 F 29.8
2:00 271.6 F 28.9
3:00 272.6 F 29.7
4:00 271.8 F 29.1

Min 271.5 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 227.2 F 0.0
10:00 227.6 F 0.0
15:00 225.8 F 0.0
20:00 223.8 F 0.0
25:00 223.2 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 18:09
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:21
CYCLE COMPLETE

02/27/2023 05:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____