
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002653
9:36 07/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 3:22
STERILIZING 4:25
136.0 C 221kPa 4:25
136.4 C 223kPa 4:55
136.4 C 225kPa 5:25
136.0 C 221kPa 5:55
136.2 C 223kPa 6:25
136.4 C 225kPa 6:55
136.3 C 225kPa 7:25
136.1 C 225kPa 7:55
136.1 C 225kPa 7:55
VENTING 7:56
AIR DRYING 8:51
CYCLE COMPLETE 9:21

Digital Signature #
09237E51E170D7A2
