

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
2.0uS / 1.2ppm  
CYCLE NUMBER 000036  
16:04 10/08/2020  
WRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 1:23  
PRESSURIZING 8:06  
STERILIZING 9:06  
136.5 C 227kPa 9:06  
136.2 C 225kPa 9:36  
136.2 C 225kPa 10:06  
136.4 C 227kPa 10:36  
136.5 C 227kPa 11:07  
136.3 C 226kPa 11:37  
136.3 C 227kPa 12:07  
136.4 C 225kPa 12:37  
136.3 C 226kPa 13:07  
136.2 C 227kPa 13:37  
136.5 C 228kPa 14:07  
136.4 C 227kPa 14:37  
136.4 C 226kPa 15:06  
VENTING 15:06  
AIR DRYING 16:00  
CYCLE COMPLETE 18:00  
-----

Digital Signature #  
6C023ACC1D37CAA7  
-----