
Midmark M11 - v1.0.5
Total Cycles: 1386

Name: RITTER2
Sterilizer ID: V2313228
08/09/2022 02:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.7 F 0.0
2:00 196.0 F 3.7
4:00 207.5 F 9.3
6:00 224.7 F 7.9
8:00 235.0 F 11.6
10:00 236.7 F 10.0
12:00 236.4 F 8.9
14:00 251.0 F 16.3
16:00 262.9 F 23.2
18:00 272.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.4
2:00 271.7 F 29.3
3:00 272.3 F 29.3
4:00 271.6 F 29.2

Min 271.3 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 220.7 F 0.0
10:00 230.1 F 0.0
15:00 226.5 F 0.0
20:00 224.8 F 0.0
25:00 225.1 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 18:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:46
CYCLE COMPLETE

08/09/2022 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____