
Midmark M11 - v1.0.0
Total Cycles: 3212

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2023 12:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.8 F 0.0
2:00 189.9 F 3.3
4:00 203.8 F 8.3
6:00 221.8 F 11.2
8:00 232.0 F 9.0
10:00 237.3 F 10.2
12:00 242.9 F 12.2
14:00 241.0 F 10.8
16:00 254.3 F 17.5
18:00 265.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.7
2:00 272.0 F 28.8
3:00 272.5 F 28.6
4:00 271.8 F 28.7

Min 271.4 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 238.7 F 0.0
10:00 232.6 F 0.0
15:00 231.7 F 0.0
20:00 226.0 F 0.0
25:00 234.3 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

05/10/2023 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____