
Midmark M11 - v1.0.5
Total Cycles: 4265

Name: RITTER2
Sterilizer ID: V2313228
01/15/2025 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 86.3 F 0.0
4:00 75.4 F 0.0
6:00 99.6 F 0.1
8:00 137.5 F 0.3
10:00 166.4 F 0.6
12:00 192.4 F 1.8
14:00 213.1 F 5.2
16:00 226.1 F 8.9
18:00 238.7 F 12.9
20:00 233.9 F 8.6
22:00 244.6 F 12.9
24:00 242.3 F 11.7
26:00 239.9 F 10.5
28:00 253.2 F 17.0
30:00 264.1 F 23.5
32:00 271.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 28.9
2:00 272.4 F 29.2
3:00 272.0 F 28.4
4:00 272.8 F 29.2

Min 271.6 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 230.5 F 0.0
10:00 177.9 F 0.0
15:00 174.4 F 0.0
20:00 170.3 F 0.0
25:00 169.5 F 0.0
30:00 170.2 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:04
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:20
CYCLE COMPLETE

01/15/2025 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____