
Midmark M11 - v1.0.0
Total Cycles: 998

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2021 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.4 F 0.0
2:00 158.9 F 2.1
4:00 175.0 F 4.6
6:00 194.8 F 8.3
8:00 217.0 F 12.9
10:00 226.7 F 7.0
12:00 237.5 F 11.7
14:00 235.1 F 8.7
16:00 241.9 F 11.0
18:00 246.3 F 13.2
20:00 258.1 F 19.6
22:00 268.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.7 F 28.9
2:00 271.6 F 28.1
3:00 272.6 F 28.8
4:00 271.5 F 28.1

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 230.3 F 0.0
10:00 220.1 F 0.0
15:00 205.6 F 0.0
20:00 220.4 F 0.0
25:00 225.1 F 0.0
30:00 189.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 23:01
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

07/12/2021 04:02 PM
ACCEPTED: X
REJECTED:
Operator: U01