
Midmark M11 - v1.0.5
Total Cycles: 2596

Name: RITTER2
Sterilizer ID: V2313228
06/30/2023 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.7 F 0.0
2:00 132.2 F 0.0
4:00 156.3 F 0.9
6:00 181.6 F 5.0
8:00 201.2 F 9.2
10:00 214.8 F 10.6
12:00 225.3 F 8.1
14:00 236.8 F 12.9
16:00 237.5 F 10.4
18:00 236.6 F 8.8
20:00 249.5 F 15.4
22:00 260.8 F 21.8
24:00 271.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.8 F 29.5
2:00 271.7 F 28.8
3:00 272.9 F 29.7
4:00 272.1 F 29.1

Min 271.7 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 198.9 F 0.0
10:00 209.5 F 0.0
15:00 218.3 F 0.0
20:00 221.0 F 0.0
25:00 223.5 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:42
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:02:59
CYCLE COMPLETE

06/30/2023 02:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____