
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000793
10:17 30/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:46
STERILIZING 3:50
136.4 C 232kPa 3:50
136.4 C 233kPa 4:20
136.1 C 230kPa 4:50
136.1 C 229kPa 5:20
136.2 C 230kPa 5:50
136.4 C 232kPa 6:20
136.3 C 232kPa 6:50
136.3 C 232kPa 7:20
136.3 C 232kPa 7:20
VENTING 7:20
AIR DRYING 8:14
CYCLE COMPLETE 8:30

Digital Signature #
16CB1887BA31F2A1
