

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3934

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2024 09:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 77.8 F 0.0  
4:00 89.0 F 0.0  
6:00 123.4 F 0.1  
8:00 159.9 F 0.2  
10:00 170.8 F 1.5  
12:00 192.2 F 4.7  
14:00 209.8 F 8.2  
16:00 221.2 F 12.1  
18:00 226.1 F 6.5  
20:00 236.1 F 11.1  
22:00 233.1 F 7.8  
24:00 245.4 F 13.3  
26:00 243.8 F 12.3  
28:00 255.3 F 18.3  
30:00 265.6 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.4 F 29.0  
2:00 271.9 F 29.1  
3:00 272.1 F 28.7  
4:00 272.4 F 29.4

Min 271.5 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.8 F 0.0  
5:00 234.8 F 0.0  
10:00 226.5 F 0.0  
15:00 217.4 F 0.0  
20:00 222.2 F 0.0  
25:00 228.8 F 0.0  
30:00 212.7 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:50  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:09:58  
CYCLE COMPLETE

09/12/2024 10:14 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_