
Midmark M11 - v1.0.5
Total Cycles: 3754

Name: RITTER2
Sterilizer ID: V2313228
07/05/2024 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.9 F 0.0
2:00 71.5 F 0.0
4:00 103.2 F 0.0
6:00 129.4 F 0.5
8:00 136.0 F 3.2
10:00 166.2 F 6.0
12:00 184.7 F 9.1
14:00 195.0 F 13.4
16:00 213.5 F 5.9
18:00 225.6 F 9.6
20:00 234.3 F 11.8
22:00 235.3 F 10.2
24:00 237.6 F 9.1
26:00 246.0 F 13.9
28:00 257.3 F 20.0
30:00 267.5 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.2 F 29.3
2:00 272.5 F 29.8
3:00 272.1 F 29.9
4:00 271.5 F 29.2

Min 271.5 F 29.1
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 213.2 F 0.0
10:00 215.1 F 0.0
15:00 206.4 F 0.0
20:00 205.1 F 0.0
25:00 207.3 F 0.0
30:00 208.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:21
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

07/05/2024 09:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____