
Midmark M11 - v1.0.5
Total Cycles: 4418

Name: RITTER2
Sterilizer ID: V2313228
03/10/2025 05:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.1 F 0.0
2:00 120.5 F 1.3
4:00 138.6 F 4.4
6:00 170.8 F 9.1
8:00 196.6 F 6.6
10:00 215.4 F 6.4
12:00 229.7 F 11.1
14:00 230.2 F 7.7
16:00 243.4 F 12.7
18:00 243.1 F 12.4
20:00 256.1 F 19.2
22:00 267.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.7 F 29.5
2:00 272.4 F 29.7
3:00 272.4 F 29.2
4:00 272.8 F 29.9

Min 271.7 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 187.1 F 0.0
10:00 199.8 F 0.0
15:00 199.5 F 0.0
20:00 202.6 F 0.0
25:00 202.6 F 0.0
30:00 202.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:19
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

03/10/2025 06:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____