
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002008
15:29 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:02
STERILIZING 4:05
136.0 C 220kPa 4:05
136.0 C 223kPa 4:35
136.2 C 225kPa 5:05
136.4 C 222kPa 5:35
136.3 C 225kPa 6:05
136.4 C 225kPa 6:35
136.4 C 224kPa 7:05
136.3 C 224kPa 7:35
136.3 C 224kPa 7:35
VENTING 7:35
AIR DRYING 8:30
CYCLE COMPLETE 9:06

Digital Signature #
8E243407BBCA8723
