
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004366
9:47 15/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:24
STERILIZING 3:18
136.0 C 220kPa 3:18
136.3 C 223kPa 3:48
136.3 C 223kPa 4:18
136.4 C 225kPa 4:48
136.4 C 223kPa 5:18
136.4 C 224kPa 5:48
136.1 C 221kPa 6:18
136.1 C 222kPa 6:48
136.1 C 222kPa 6:48
VENTING 6:48
AIR DRYING 7:39
CYCLE COMPLETE 8:32

Digital Signature #
507CD9B3AA651418
